

Power requirement:

System operates between 200-240 AC, 50 Hz. Maximum Power consumption is 400 VA only.

Required Accessories:

Broadband (2-7MHz) convex probe with tissue harmonic

Broadband (4-13MHz) Linear probe with tissue harmonic

Suitable online UPS should be supplied

B/W Thermal Printer should be supplied

Quote Optional

Broadband (3-9 MHz) TV/TR probe with tissue harmonic

Broadband (1-4 MHz) Phased Array cardiac probe with tissue harmonic

1. Unit should be / CE approved all radiation protection certifications.
2. If the supplier is OEM reference will be given for better after sales and AMC condition.
3. The firm should have high repute in National / International market for more then ten years.
4. The firm should give the name of institutions where similar unit has been supplied in last three years.
5. The firm should give the copy of purchase order for justification of cost in which previous unit is supplied.
6. Name and address of local service center.

Dr P.K.Mittal