

T.No. 65.12.8032

Specifications for CNS System for ICU.

Major parameters.

Connectivity to single Para, three Para or five Para monitor.

Input connection from 8- 16 beds.

On line monitoring of all diagnostic parameters.

It should have ST segment analysis and arrhythmia analysis.

It should have 96 hours of Trend memory

Automatic data save in case of power failure.

Zooming and freezing facility for each channel separately.

It should have Electro Surgical unit & Defibrillator Protection.

Minor parameters.

It must be of an international repute with a standardization certification i.e. CE certified or equivalent.

The company should submit the list of major hospitals where same kind of equipment has been installed in last three years, and also submit a satisfactory certificate of the same.

The firm should have an authorized service centre in Jabalpur or should clearly mention about the service centre, which can provide the service within 24 hours of registering a complaint.

Comprehensive AMC for 7 yrs after 2 years warranty.