

Annexure to Tender No.65125007/5008

1. **Item:** - Fuel cell sensor based breath alcohol analyzer.

Requisition No:- 63289736 Dated :- 04.05.2012

Fuel cell sensor based breath alcohol analyzers should follow the following specifications-

S.N	Description	Details
1.	Specification No	RDSO/SPN/TR/290/BA (Rev,0) -2003 issue date 28.04.2003 (Copy attached) or latest if issued by the RDSO.
2.	Features required	It should be compatible to attach with computer for report data transfer and saving for record.
3.	Printer and it's features required	<p>Separate Printer must be supplied (Not inbuilt) and it Shall have the facility to connect with Breath Analyzers for taking printout of test data.</p> <p>Working on plain paper roll and easily available ribbon cartridge.</p> <p>Print data should not fade away with time.</p> <p>Print out shall give details of date and time or blank & subject tests, test number, blank and subject test readings along with spaces for signature and details of subject, operator and witness (or customized as required by the consignee)</p> <p>Printer shall have auto shut down facility after printing to save power.</p>
4	Model	Hand held type
5	Type of cover	Leather case
6	Battery charger cum adapter	Built in or separate
7	Item to be supplied with each set	<p>1.Nozzle- atleast-25 No. (If required as per design)</p> <p>2.Operating and trouble shooting manual.</p> <p>3.Quality assurance manual as per item 6.72</p>