

Technical Specification of COAGULOMETER FOR CENTRAL HOSPITAL JBP.

Sr. No.	Coagulometer	Specification
01	Principle	Scattered light detection
02	Light Source	LED
03	Parameter	PT, PTT, APTT, FBG, TT, NT, Extrinsic factor-(II, V, VII, X) Intrinsic factor- VIII, IX, XI, XII
04	Detection Range	Upto 470nm.
05	Temp. Control	37°C ± 0.5°C programmable incubation
06	Power Consumption	< 100w
07	Voltage	85-250V
08	Printer	Built in printer can be connected to extension printer.
09	Memory	Can save more than 100 results.
10	Pipetting	Automatic measuring pipetting.
11	Test channel	Each channel is independent same or different. It can be tested simultaneously.

/MD/WCR/JBP.