

5

Technical specification for Procurement of Defibrillator with external pace maker with recorder.

- 1) The defibrillator should have biphasic technology having energy selection of 1-200 joules.
- 2) The machine should have facility for ECG monitoring, defibrillation, external pacing & recorder.
- 3) It should be capable of monitoring ECG through ECG Cables, electrodes & paddles.
- 4) The machine should have defibrillation facility for adult & pediatric patients.
- 5) The machine should have ECG waveform display on bright screen along with other vital numeric information.
- 6) The machine should be compact, portable with built in rechargeable battery & light weight.
- 7) The machine should have inbuilt recorder for printing ECG trace & stored information.
- 8) The machine should have facility of up gradation for vital parameter such as NIBP & SPO2.
- 9) The machine should have user selectable alarm setting.
- 10) The machine should work on mains (without battery) and on battery as well.
- 11) The machine should have AED feature as inbuilt with manual override for manual operations.
- 12) The machine must be supplied with all the essential accessories in 2 set & moveable trolley.
- 13) Service centre must be in Jabalpur (M.P.) & CMC for more than 2 years.
- 14) Machine should have warranty of 2 years.

*basu/bs/ls*  
DMO (Phy)

~~Add CHD (M&P)~~

~~MD /CH /WCR~~

~~CMD /WCR~~