

22
9

SPECIFICATIONS

- Weight - 10 – 20 Kg approx.
- Display screen - Colour TFT LCD.
- Printer - Thermal line printer with auto cutter.
- Chin rest - Motorized.
- Power saving mode - Yes.
- Alignment - Fully automatic 3D alignment.
- Measurement mode - Fully Automatic/Automatic/ Manual
- IOP measurement range - 0 – 60 mmHg. (Approx.)
- Measurement range - Air Puff non-contact method.
- Measurement principle - 11 mm (approx)
- Working distance

Pachymeter Measurement

- Measurement range - 150 – 1300 microns (approx)
- Measurement principle - Quantized slit image on central cornea.

Essential accessories :-

- Motorized table for keeping the machine.

Ajay Dogra
(Dr.Ajay Dogra)
Sr.DMO(Ophthal) JBP.

Amir Chand

अमीर चन्द
AMIR CHAND
उप मुख्य सामग्री प्रबंधक/डीजल
Dy Chief Matenal Manager/DSL
पश्चिम मध्य रेल/जबलपुर
West Central Railway, JBP.

000019

Annexure-II

To be submitted by the firm along with offer

39

Tender specification as per Annexure-I	Specification quoted by the firm	Deviation if any with clarification

Signature of firm's