

(7)

Technical Specification for Digital Holter Analyzer Machine

- 1) The Digital Holter Recorder must be of 12 channel mode.
- 2) The Digital Holter Recorder should have 5 lead arrhythmia detection facilities.
- 3) Power supply should be AC (100-250 V @ 50/60 Hz)
- 4) The Digital Holter Recorder should have 24 hour recoding facility.
- 5) Battery must be built-in rechargeable Li-ion Battery or External Battery with charger.
- 6) The machine should be compact, portable & light weight.
- 7) Digital Holter Analyzer must have facility for Heart rate Variability analysis, Arrhythmia detection facility such as VE, SVE, Bigamini, Trigemini, Tachycardia, Bradycardia, R on T, Couplets, Pauses, ST Segment Analysis facility, QRS Templates analysis.
- 8) Data Storage must be in inbuilt & data transfer through USB.
- 9) Digital Holter Analyzer must have complete editing facility for Arrhythmia, HR, and ST Segment. Etc.
- 10) Digital Holter Recorder must have event marker button, Beep Buzzer, Alarm Switch.
- 11) Digital Holter analysis system must be computerized (Windows based) along with facility of Colour laser printing of reports and must have at least 2 hour battery backup via UPS. All these things must be branded.
- 12) Holter analyzer must have facility of editing.
- 13) Holter analyzer must have facility of PDF file for email attachment.
- 14) Patient cable must be detachable.
- 15) Digital Holter Analyzer must be with unlimited data storage in Hard disk & also with 32 GB, USB memory stick.
- 16) The machine must be supplied with all the essential accessories in 2 set like patient cable, recorder cover with belt, USB data cable & 1 table for PC and one carry bag.
- 17) Service centre must be in Jabalpur (M.P.) & CMC for more than 2 years.
- 18) Machine should have warranty /Guarantee of 2 years.

Minor Parameter

- 1) Firm may submit its past supply performance/ credential.
- 2) The company should submit the list of major hospitals where same kind of equipments has been installed in last three years.
- 3) The firm should clearly mention address & phone number of service centre, which can provide the service within 24 hours of registering complaints.
- 4) The firm must confirm for entering in to AMC/CMC with consignee after expiry period.

(10)

13/01/14

Bankur
DMO (Phy)

Annexure-II

To be submitted by the firm along with offer-

Tender specification as per Annexure-I	Specification quoted by the firm	Deviation if any with clarification

Signature of firm's

cyg
13/1/14