



**TECHNICAL SPECIFICATION FOR TEST BENCH FOR CALIBRATION OF  
PRESSURE GAUGES RANGE - 0 TO 35 KG/CM<sup>2</sup> & 0 TO 500 PSI  
LEAST COUNT 0.025 KG/CM<sup>2</sup> & 0.5 PSI**

9

Major parameter	Test bench suitable for calibration of Pressure Gauges Range - 0 to 35 Kg/cm <sup>2</sup> & 0 to 500 PSI least count 0.025 kg/cm <sup>2</sup> & 0.5 PSI. As per Drawing attached 1. Length - 54 inch (± 6 inch) 2. Width - 30 inch (± 3 inch) 3. Height - 60 inch (± 6 inch) 4. Total gauges tested at a time - 4 nos. + 1no. Master gauge 5. Diameter of gauge's - 2" to 8"
Minor parameter	Nil
Concomitant accessories	Nil
Optional ; accessories	Nil
Spares	As per recommended by Firm
AMC requirement	Nil
Warrantee requirement	24 months from the date of machine receipt.
Special requirement	Two copies of printed original operation & maintenance procedure manual literature drawing circuit diagram to be provided in binding form.

  
SSE (Air Brake)

  
ADME (D) ET  
एच. एन. सिंह  
H. N. SINGH  
वरि.मं.यां.इंजी. (डी) इटारसी प.म.रेल  
A.D.M.E. (D) Itarsi W.C.Rly.

  
Sr.DME (D) ET  
एस. के. सिंह  
S. K. SINGH  
वरि.मं.यां.इंजी. (डी) इटारसी प.म.रेल  
Sr. D.M.E. (D) Itarsi W.C.Rly.

अनूप सिंह यादव  
Anoop Singh Yadav  
वरिष्ठ सामग्री प्रबंधक/सामान्य एवं निर्माण  
Senior Material Manager/G&C  
प.म.रे., जबलपुर/W.C.R., JBP.

**Annexure-II**

**To be submitted by the firm along with offer-**

Tender specification as per Annexure-I	Specification quoted by the firm	Deviation if any with clarification

**Signature of firm's**