

(178)

Specification of Multi Para Monitor for Divisional Hospital, Kota

Major parameters : No deviation (Tolerance) accepted.

1. Machine should be CE mark and as per technical standard EN 6060 1-2-4 with cautery filter.
2. Standard configuration of
 - i) ECG
 - ii) Respiration
 - iii) SPO2
 - iv) NIBP
 - v) Temp.
3. Drug dose calculation / HRV analysis / OXY CRG.
4. 3 - lead / 5 lead monitoring
5. Alarm event review.
6. Pacemaker Detection Function.
7. Built in three channel thermal recorder (optional)

Minor Parameters :

- Battery good quality mark with backup atleast 2 hrs.
 - Accessories : All accessories must be international level quality mark.
 - Warranty period : 02 Years.
 - AMC / PME : 5 Years after warranty period.
 - i) Active colour TFT Display Touch Screen.
 - ii) Waveform : 7Channels
 - iii) 96 hours of trend
 - iv) Battery backup
 - v) Arrhythmia & ST analysis
 - vi) Call should be attended within 48 hrs.
 - vii) Preventive check and maintenance should be carried out.
8. Terms & Condition : Installation at operation theatre of Divisional Railway Hospital, Kota.
9. Rates of AMC should be quoted likely source of suppliers.
- i) M/s Vediotronics
410, Shopping Centre, Kota
 - ii) M/s Durga Surgical
8-B, Gulab Bari, Kota
 - iii) M/s Shreya Medical
Banta-16, Vigyanagar, Kota

अनूप सिंह यादव

Anoop Singh Yadav

वरिष्ठ सामग्री प्रबंधक/सामान्य एवं निर्माण

Senior Material Manager/G&C

प.न.रे., जयलपुर/W.C.R., JBP

Usha Rani
14/3/14
(Dr. Usha Rani)
ACMS (Ansth.) Kota

Annexure-II

To be submitted by the firm along with offer-

Tender specification as per Annexure-I	Specification quoted by the firm	Deviation if any with clarification

Signature of firm's

