

(29) (28)

TECHNICAL SPECIFICATION OF OPERATIVE HYSTEROSCOPY SET

S.No.	Description	Qty.
1.	Forward-Oblique Tele-scope 30o, mm, length 30 cm, autoclavable, fibre optic light transmission incorporated color code : red	1
FOR DIAGNOSTIC HYSTEROSCOPY		
2.	Examination Sheath, o 5.2 mm, with 1 Luer-lock adaptor, for use with CF-Examination sheath.	1
3.	Continuous-Flow Examination Sheath, 6.2mm, with 1 LUER-lock Adaptor, for use with Examination sheath.	1
FOR OPERATIVE HYSTEROSCOPY		
4.	Operating Sheath, size 5.4 mm, with channel for semirigid 5 Fr. Operating instruments, with 1 stopcock and 1 LUER-Lock adaptor, for with CF operating sheath.	1
5.	Continuous-Flow operating Sheath, size 6 mm, with 1 stopcock and 1 LUER-Lock adaptor, for use with operating sheath.	1
6.	Scissors, semirigid, blunt, single action jaws, 5Fr., length 34 cm.	1
7.	Biopsy and Grasping Forceps, semirigid, double action jaws, Fr., Length 34 cm	1
8.	Punch, through-cutting, single action jaws, semirigid, 5 Fr., length 34 cm.	1
9.	Polypectomy Loop, unipolar, 5FR., length 34 cm.	1
10.	Bipolar dissection electrode, semirigid, 5 Charr., length 36 cm.	1
11.	Needle Electrode, unipolar 5 Fr., length 34 cm	1
12.	Tenaculum grasping Forceps, double action jaws, semirigid, 5 Fr., length 34 cm.	1
CO2 ENDOFLATOR		
S.No.	Description	Qty.
13.	Electronic Endoflator set, SCB operating voltage : 100-240 VAC, 50/60 Hz Consisting of : Electronic Endoflator, with integrated SCB Module 400 A Power cord Silicone tube, sterilizable Universal wrench SCB Connecting Cable Sterile filter, Package of 10 pieces.	1
14.	High Pressure Hose. American connection – pin-index connection length 55 cm	1

जे. एन. मीणा
J. S. MEENA
उप मुख्य सामग्री प्रबन्धक/जी.एण्ड सी.
Dy. Chief Materials Manager/G&C
पश्चिम मध्य रेल, जबलपुर
West Central Railway, Jabalpur

M. L. L. L.
6/3/13

Annexure-II

To be submitted by the firm along with offer-

Tender specification as per Annexure-I	Specification quoted by the firm	Deviation if any with clarification

Signature of firm's