

## Ophthalmic Surgical Microscope

### Technical Specification

1. Surgical Microscope - CE certified
2. Eye Piece - 10 X
3. Binocular Tube - 45 degree inclined /90 degree, Tilttable F= 170 mm
4. Magnification changer- Manual Step. (Option Motorized step zoom 1:6)
5. Magnification - 0.4 X, 0.6X, 1.0 X, 1.6 X, 2.5 X
6. Objective - F= 200mm (Option F=175 mm)
7. Focusing - Motorized Focusing
8. Stand - Mobile floor stand on four wheels
9. Weight - 80 Kgs- 120 Kgs
10. Balancing - Counter balanced suspension arms.

### Electrical Specification

1. Power Supply - -/10- 230V, 50-60 Hz
2. Light source - LED

### Illumination

1. Bright LED
2. Yellow, Retinal guard and integrated heat filter.
3. X-Y Coupling.
4. Push Button for initial position of X-Y

### Standard Accessories

1. Eye piece 12.5 X
2. Anterior Segment Imaging and video documentation. Beam splitter, C-mount, CCD Camera, in built Camera (Optional)

Amnepune-I

12

3. Sterilizable caps- two sets
4. Dust cover - 1

CMC/AMC required as per ~~TRBPS~~ conditions.

*Aj 14/6/16*  
(Dr Ajay Dogra)  
ACHD (Ophth) CH JBP

## Annexure-II

To be submitted by the firm along with offer-

Tender specification (As per Annexure-I)	Specification quoted by the firm	Deviation if any with clarification

To be submitted by firm along with offer:-

Supplier Name:-	
Basic Rate( in Rs):-	
ED:-	
CST:-	
Freight:-	
Other charges:-	
All inclusive Rate of machine:-	
<b>CMC Charges:-</b> For 1 <sup>st</sup> Year after expiry of warranty. For 2 <sup>nd</sup> Years after expiry of warranty. For 3 <sup>rd</sup> Years after expiry of warranty. For 4 <sup>th</sup> Years after expiry of warranty. For 5 <sup>th</sup> Years after expiry of warranty.	
Total Cost (Inclusive of CMC charges):-	

Signature of the firm