

No:- 65-17-5012

**Specification of Orthopaedic Operation Table under M&P programme for the Year  
2016-2017 under GM's power for DRH BPL**

S.N.	Category No.	Description
1	Major Parameter	<ul style="list-style-type: none"> <li>• Mobile C-Arm Table Eccentric Type pillar base, Electro Hydraulic Table, Extremely smooth, precise, jerk-free movements of operation theatre table with just a touch of button.</li> <li>• <b>Table Top</b>-composite imported radiolucent material for low radiation and stain free.</li> <li>• <b>Length</b>- 1676 mm (+ 300mm detachable Head Rest)</li> <li>• <b>Width</b>-482 mm</li> <li>• <b>Side Rails</b>-736 mm x 10 mm stainless side rails fitted on both sides.</li> <li>• <b>Table Movements-</b> <ul style="list-style-type: none"> <li>(a) <b>Vertical</b> – 810 mm min to 1065 mm max. Remote controlled</li> <li>(b) <b>Trendelenberg &amp; Rev.Tr.</b>- 35 ° Remote controlled</li> <li>(c) <b>Lateral Tilt</b>- 20 ° left &amp; right Remote controlled</li> <li>(d) <b>Back Section</b>- Flex &amp; Reflex (75 ° up &amp; 45 ° down) Remote controlled</li> <li>(e) <b>Autolevel</b>- Zeropositioning of table top (Trendelenberg) Remote controlled</li> </ul> </li> <li>• <b>Attachments-</b> <ul style="list-style-type: none"> <li>(a) Head Rest- Detachable</li> <li>(b) Lower Limb Fracture Att. Detachable</li> </ul> </li> <li>• <b>Table Controls-</b> <ul style="list-style-type: none"> <li>(a) Drive System --- Non Hydraulic- leak proof design requiring negligible maintenance.</li> <li>(b) Remote Control - Corded remote with control buttons.</li> <li>(c) Mobility- Fitted with castor wheels for free table mobility &amp; floor lock.</li> <li>(d) Dual Control Console- Unique dual control console on the table itself, so that it can be Operated even without the remote control.</li> </ul> </li> <li>• Maximum Patient Load Capacity- 130 to 150 kgs.</li> <li>• Standards Accessories: Shoulder supports-2 nos, Body Side supports-2 nos, Arm rests-2 nos., Patient Body Strap-1 no., Mattress.</li> </ul>

		<ul style="list-style-type: none"><li>• <b>Neurosurgery and Orthopaedic Attachments</b></li><li>(a) Laminectomy frame for spine.</li><li>(b) Cervical Spine Horse shoe attachment</li><li>(c) Femur in Lateral position attachment</li><li>(d) Tibia Nailing attachment</li><li>(e) Shoulder Arthroscopy attachment in sitting position</li><li>(f) Knee Arthroscopy attachment</li><li>(g) 2 section Arm surgery attachment</li></ul> Should be CE/IS 13485 certified
2	<b>Minor Parameter</b>	NIL
3	<b>Concomitant Accessories</b> <b>Concomitant Accessories</b>	NIL
4	<b>Optional Accessories</b>	NIL
5	<b>Spares</b>	NIL
6	<b>AMC/ CMC Requirement</b>	NIL
7	<b>Warranted Requirement</b>	Warranty one year from the date of installation.
8	<b>Special Requirement</b>	NIL

  
22/6/17  
CMS/BPL